

Introduction to the new calibration sources and their performance

EMPIR 19ENV01 traceRadon

This project 19ENV01 traceRadon has received funding from the EMPIR programme co-financed by the Participating States and from the European Union's Horizon 2020 research and innovation programme.

19ENV01 traceRadon denotes the EMPIR project reference.

Petr Otahal, Eliska Fialova – SÚJCHBO, v.v.i. Activities A1.3.3 & A1.3.4





Low-level Rn-222 emanating sources





- ❑ Radon measurement techniques are simple, efficient and precise.
- ❑ Levels of relevant activity concentration in European dwellings are laid down ($300 \text{ Bq}\cdot\text{m}^{-3}$).
- ❑ Developing and improving of calibration procedures is still actual.




Low-level Rn-222 emanating sources

- ❑ Stainless steel cylindrical case, ball valves.
- ❑ Steel tray with Ra-226 placed in the middle of this cylindrical case - radon releases from this thin layer.
- ❑ Flow-through mode.
- ❑ The emanation coefficient was determined by measuring the activity of the RnDP (Pb-214/Bi-214) - the activity of Ra-226 is almost equal to 1.
- ❑ The detection
- ❑ efficiency of the gamma photons was calculated by the MCNP code (Monte Carlo N-Particle Transport).



- 1 – ball valve
- 2 – emanator
- 3 – holder
- 4 – flange
- 5 – retaining screw

Material: Dural, brass, stainless steel, Teflon, epoxy resin





Low-level Rn-222 emanating sources



Český metrologický institut

Okružní 31, 638 00 Brno
tel. +420 545 555 111
www.cmi.cz

Pracoviště: Oblastní inspektorát Praha, Radiová 1136/3, 102 00 Praha 10
Oddělení výroby standardních radionuklidových zdrojů, Radiová 1288/1a
tel. +420 266 020 460

CERTIFIKÁT etalonu aktivity

Číslo certifikátu: 1035-SE-40175-21 **Typ:** RF1 **Výrobní číslo:** 230321-221458

Radionuklid: Ra-226

Poločas Ra-226: 1600 (7) let

Aktivita: 1,136 (17) kBq

Poločas Rn-222: 3,8232 (8) dní

Radonová výdejnost: 0,0023 (1) Bq/s

Emanační schopnost: 0,9552 (19)

Radionuklidové nečistoty: -


Vnitřní objem mezi ventily: 200 (5) cm³

Referenční datum: 1. 4. 2021






Low-level Rn-222 emanating sources – background test



Low-level Rn-222 emanating sources – laboratory calibration

Achieving of low-level radon activity concentration:

- Constant dosing of radon.
- Defined ventilation.
- Radon free air.





Low-level Rn-222 emanating sources – The reference level of radon

□ Model of constant radon input and constant ventilation:


$$a(t) = a_0 \cdot e^{-(\lambda+k).t} + \frac{R}{V(k + \lambda)} (1 - e^{-(\lambda+k).t})$$

$a(t)$ radon activity concentration in time t ($\text{Bq}\cdot\text{m}^{-3}$)
 a_0 radon activity concentration in time zero ($\text{Bq}\cdot\text{m}^{-3}$)
 λ radon decay constant (h^{-1})
 k air exchange intensity (h^{-1})
 t time (h)
 R radon input rate ($\text{Bq}\cdot\text{h}^{-1}$)
 V volume of radon chamber (m^3)






Low-level Rn-222 emanating sources





Low-level Rn-222 emanating sources







Low-level Rn-222 emanating sources – field calibration



Radon activity concentration 1 – 20 Bq·m⁻³

Maximal flow rate: 10 L/min

Only for devices in flow measuring regime




Inversion weather

necessary to know the background activity of radon monitor





Low-level Rn-222 emanating sources – field calibration








Low-level Rn-222 emanating sources – field calibration



Radon activity concentration, two AlphaGuards



Radon activity concentration, one AlphaGuard





Crucial parameters – Field calibration




- The background activity of the measuring device has to be known (should be analyzed in the radon free atmosphere, no more than $10 \text{ Bq}\cdot\text{m}^{-3}$),
- have to be checked the flow rate value of the measuring device (the SUJCHBO AlphaGuard was checked by Gilibrator-2 – calibrated by the CMI),
- calibration should be performed at the lowest radon concentration in the outdoor atmosphere (8:00 a.m. – 6:00 p.m.),
- in the case of using a CMI flow source, it is necessary to replace the atmosphere in the source at least 100 times during minimally 30 minutes and immediately use the source for calibration.





Crucial parameters – Field calibration



The result is $120.1 \pm 27.9 \text{ Bq}\cdot\text{m}^{-3}$ after subtracting the background activity and outdoor radon concentration from the average results, representing the time when the radon source is connected to the AlphaGuard.

Parameter	Value
Background activity	$2 \text{ Bq}\cdot\text{m}^{-3}$
Set flow rate value	1 L/min
Measured flow rate value	1,156 L/min
Source radon activity concentration	$119,4 \pm 6 \text{ Bq}\cdot\text{m}^{-3}$





Field calibration

